

QCA – Biochemical Calibrator**ΠΙΝΑΚΑΣ ΟΡΟΥ ΤΙΤΛΟΠΟΙΗΣΗΣ****ΤΙΜΕΣ ΑΝΑΦΟΡΑΣ****LOT: 240110****Exp.: 2 / 2027**

	ΤΙΤΛΟΣ	ΑΠΟΚΛΙΣΗ	ΤΙΜΗ ΣΤΟΧΟΣ	ΜΟΝΑΔΕΣ
	ALBUMIN	+/- 0.278	2.39	g/dl
	Alk. Phosphatase DGKC	+/- 36.5	391	U/L
	Alk. Phosphatase IFCC	+/- 26.5	282	
	α-Amylase	+/- 19.5	202	
	BILIRUBIN Total	+/- 0.262	3.61	mg/dl
	BILIRUBIN Direct	+/- 0.151	1.01	
	CALCIUM	+/- 0.740	9.50	
	CHOLESTEROL	+/- 15.6	210	
	CREATININE	+/- 0.494	3.00	
	GLUCOSE	+/- 13.8	198	
	CPK	+/- 26.5	472	U/L
	GAMMA GT	+/- 16.7	123	
	GOT (ASAT)	+/- 13.8	110	
	GPT (ALAT)	+/- 8.3	106	
	HDL	+/- 10.7	68.0	mg/dl
	IRON (Ferene)	+/- 19.2	146	μg/dl
	LDH	+/- 64.9	556	U/L
	MAGNESIUM	+/- 0.469	3.38	mg/dl
	PHOSPHORUS	+/- 0.350	4.99	
	Total PROTEIN	+/- 0.415	3.27	g/dl
	TRIGLYCERIDES	+/- 11.9	190	mg/dl
	UREA	+/- 9.10	81.6	
	URIC ACID	+/- 0.877	6.68	